



क्रय अनुभाग
वन अनुसंधान संस्थान
देहरादून-२४८००६
GSTIN 05AAAAI1708G2Z6
PAN NO. AAAAI1708G

Purchase Section
Forest Research Institute,
DEHRADUN-248006
Phone: 0135-2224205
http://fri.icfre.gov.in/tender/
Email: po_fri@icfre.org



Tender Reference No. X-7/POFRI/CHE/VKV/AICRP/2020-21

Date 7th August, 2020

Tender ID: 2020_ICFRE_573020_1, Website: <https://eprocure.gov.in/eprocure/app>

CORRIGENDUM

Name of item: Ultra Performance Liquid Chromatograph- Quadrupole Time-of-Flight Mass Spectrometer (UPLC-QTOFMS)

After pre-bid meeting, following amendment has been made regarding above mentioned purchase/tender:

1. The payment term will be 80% against satisfactory supply and installation and 20% after satisfactory performance of the equipment and training to the FRI scientists by supplier free of cost and on submission of bank guarantee for equivalent amount i.e. 5% cost of order with validity upto warranty period of five years from the date of satisfactory installation / commissioning and acceptance of supply. No other payment terms will be considered. It must be clearly mentioned in the offer whether our payment term is acceptable to the bidder.

2. WEIGHATGE SYSTEM should now we read as follows:

SI.No.	Name & Required Specifications of Item(s)	Total Marks
		70
		Weightage
1.	MS/MS Sensitivity	Max ^m Marks 5
1.1	500:1 to 1499:1	3
1.2	1500:1 or higher	5
2.	Resolution upto m/z 3000	Max ^m Marks 22
2.1	40000 FWHM-50000FWHM	13
2.2	>50000 FWHM	22
3.	Mass Accuracy (in ppm)	Max ^m Marks 15
3.1	2.50-2.01	5
3.2	2.00 to 1.00	10
3.3	<1.00	15
4.	Detector Acquisition Speed	Max ^m Marks 5
4.1	4 GHz	3
4.2	> 4GHz	5
5.	Work flow (Both semi targeted (using internal library) and Non targeted (using External library) screening can be performed at the same time	Max ^m Marks 5
5.1	Yes	5
5.2	No	0
6.	Software integration with Chemspider database	Max ^m Marks 5
6.1	Yes (allows for search and data reporting within a single window)	5
6.2	Yes (does not allow for search and data reporting within a single window)	2
7.	Component identification: No. of natural product compounds in database library	Max ^m Marks 8
7.1	2000 -2500	3
7.2	2501-5000	5
7.3	>5000	8
8.	Isotopic fidelity	Max ^m Marks 5
8.1	<5%	5
8.2	5% or >5%	0

The other terms and condition will remain unchanged except the above.

Purchase Officer, FRI