

Corrigendum

Tender No.X-53/2018-19/Purchase Tender ID 2018_ICFRE_423365_1

Dated 03rd Jan 2019.

Revised Specification; - For item No. 1. "Chip Based Electrophoresis System"

1. Point No. 4 of item No. 1. Chip Based Electrophoresis System may be read as; "The Microchips based on microfluidic technology must be reusable for minimum of 2000 samples and should have shelf life of minimum 6 months after opening.
2. Point No. 12 of item No. 1. Chip Based Electrophoresis System may be read as; "The time for processing the size determination of fragments from 100 to 5000 bp should "be in the range of 25 to 50 sec. per sample."

Revised Specification for item No. 4. "Automated Flash Chromatography System"

1. Point No. 3. of item No. 4. Automated Flash Chromatography System-Solvent flow Rate:- may be read as; "1 to 150 ml /min or above"
2. Point No. 4. Pumping System:- may be read as; Pumping system should be sufficient to give "15 bars/220psi or more"


Purchase Officer, FRI