



क्रय अनुभाग
वन अनुसंधान संस्थान
देहरादून-२४८००६
GSTIN 05AAAAI1708G2Z6
PAN NO. AAAAI1708G

Purchase Section
Forest Research Institute,
DEHRADUN-248006
Phone: 0135-2224205
http://fri.icfre.gov.in/tender/
Email: po_fri@icfre.org



निविदा दस्तावेज/ Tender Document

Ref. No X-67/GTI(PP)/SB&RKM/CAMPA/2019-20

Date: 12/02/2020

Physical Bids under two bid systems (Technical and Commercial) are invited from the reputed manufacturer /authorized supplier/dealer for the following items. The bids complete in all respect must reach to the Purchase Officer, Purchase Section, Forest Research Institute, P.O. New Forest, DEHRADUN-248006 Uttarakhand, India on or before:

Date	28-02-2020	Time	03:00PM
------	------------	------	---------

वस्तु की तकनीकी विशिष्टताये / Technical Specification of item:

Sr. No.	Name of item / Specifications/	Quantity	EMD (Rs.)
1.	Deep Freezer (-80deg. C)	02	44,000/-
2.	High throughout horizontal gel electrophoresis systems with power pack	01	6,000/-
3.	Thermal Cycler (PCR machine)	02	24,000/-
4.	Real Time PCR or qRT-PCR	01	30,000/-
5.	Refrigerated Micro centrifuge	01	10,000/-
6.	Refrigerated Incubator Shaker	01	6,000/-
(For Detailed Technical Specification Plz refer Annexure-VI)			
नोट-1: कृपया वस्तु का मूल्य एवं कर अलग-अलग दर्शाए / Please quote the rate & Taxes of the item separately.			
Note-2: E-way bill is to be issued by the Venders/Transporter			
Note-3: The HSN / SAC Code of the item must be mentioned.			
For Reasonability of Rates, The firm must enclose the last two supply orders for the same item with Financial Bid (Annexure-IV) only.			
Technical Clarifications:	Dr. Santan Barthwal	Genetics & Tree Improvement Division, FRI Dehradun	Ph: 0135-222-4408 Email: barthwals@icfre.org

Chapter-1: Instruction to bidders

A. The Tender should be enclosed with proper certifications like Agency Certification, Authorization certificate and /or Proprietary Certificate, as the case may be, in support of your offer.

B. बोली का जमा किया जाना / Submission of Bids:

1) The bids should be submitted in two parts-

- The Technical Bid (Annexure-I,II,III and VI with Supporting/Required documents) and
- The Financial Bid (Annexure-IV & V).

Both in separate sealed covers and then these two sealed packets (with the identification statement clearly written- Technical Bid or Financial Bid) should be put in one covering sealed envelope labeled on the left hand top corner of their envelope mentioning followings:

Tender Ref: X-67/GTI(PP)/SB&RKM/CAMPA/2019-20

Bid Submission End Date/Time: 28-02-2020 by 03:00 PM

Name (in full) and complete address of the firm should be mentioned on left hand bottom corner of the covering envelop.

- The bids should be printed on **official pad** preferably with GST Number of the firm.
- The quotations should bear full details and where possible may be duly supported with catalogues, pamphlets, literature, samples of the item/items as the case may be for comparing the quality and rates of the item(s).
- Quotation received after the closing date and time shall not be considered.
- The transportation cost, insurance charge etc., if any, Percentage/ rate of GST or all other taxes and duties should be clearly mentioned.
- The bids / rates / tenders should remain valid for a minimum period of 45 days from the date of opening.
- Since this purchase being for research purpose, the Forest Research Institute, Dehradun is registered in DSIR, New Delhi; it is therefore exempted from the payment of custom duty / GST up to the limit of exemption as per the Govt. Rules. The bidder should quote accordingly.
- There should not be any over-writing or cutting on rates quoted. However, all cutting/corrections must be duly authenticated else the offer will be rejected.
- The rates shall not be subject to escalation of any nature.

- 10) The rates quoted should be applicable to educational institutions and any cost advantage received in lieu thereof should be passed on the Institution.
- 11) While quoting/ sending rates, the firm shall give an undertaking as per Annexure-II.
- 12) The quotations should be given for the items in the same order as in the enquiry letter.

C. निविदा शुल्क/ Tender Fee एवं/and अग्रिम धन राशि / Earnest Money Deposit (EMD)

The Technical Bid should accompany Tender Fee of Rs.590/- (including 18% Service Charge) in the form of Demand Draft only and an amount of EMD mentioned at the first page of this tender document only for the equipment(s) bidding by the bidder in the form of FDR only, issued only from any nationalized bank of India drawn in favour of "Director, Forest Research Institute, Dehradun" valid 90 days beyond the validity period of bid. The bids received without Tender Fee and EMD will be rejected.

Bidders registered with MSME/NSIC/KVIC etc as per Public procurement policy for Micro & Small Enterprises (MSE) order 2012 are exempted categories from payment of Tender Fee and EMD provided that the registration certificate issued by any one of these below mentioned agencies must be valid as on close date of tender.

D. निष्पादन बैंक प्रत्याभूति / Performance Bank Guarantee (PBG):

5% of the cost price in the form of Bank Guarantee or FDR pledged in favour of "Director, Forest Research Institute, Dehradun" is required at the time of installation of the equipment from the successful bidder only which should remain valid at least during the warranty period plus sixty days.

- 1) The technical Bids will be opened as given below. Representatives of the firms (s) may be present at the time of opening the technical bid if they so desire at below schedule (No information shall be given to the late arrivals regarding the bids):

Bid Submission Opening Date/Time:	28-02-2020	04:00 PM	Venue: Purchase Section, Main Building FRI Dehradun
--	-------------------	-----------------	--

- 2) The Financial bids of only the Technically Qualified bidders will be opened. The date and time for opening the Financial Bid will be communicated to the respective firm later on by post/e-mail.
- 3) Representatives of the firms (S) may be present at the time of opening the financial bids if they so desire.

Chapter-2: Conditions of Contract

A. कार्य अनुबन्ध / Award of Contract:

The final selection of the bidder for the award of the contract will be made on the basis of the lowest commercial bid amongst the Technically Qualified Bidders.

B. पूर्व आपूर्ति निरीक्षण/ Pre-supply Inspection:

Authorized representative of the Institute shall make the final inspection before supply of the item at site of the firm, if required.

C. वस्तु की आपूर्ति / Supply of Item:

The whole supply as per order shall have to be completed within the time mentioned in the order failing which the FRI Dehradun shall have the right to accept or reject any quantity of items ordered. The firm will have to arrange for supply of the material/ equipment in good condition.

D. भुगतान की शर्तें / Payment Terms:

1) Payment in INR for supply of goods:

Where installation is not required: 100% payment will be made after satisfactory delivery of material duly certified by the HOD.

OR

Where installation is required: 100% payment will be made after satisfactory delivery and installation of the material duly certified by the HOD.

OR

100% Advance will be paid against submission of 110% Bank Guarantee of the Total Order value.

2) Payment in INR for AMC/CMC/Service:

Quarterly/Half yearly after satisfactory completion of work/service duly certified by the H.O.D./P.I.

3) Payment in Foreign Currency for supply of goods:

Payment will be made as below after deducting Indian Agency Commission (if any). IAC is payable in Indian Rupee after satisfactory installation.

- (i) Payment through Letter of Credit (LC).
- (ii) Payment 100% through Sight Draft
- (iii) Payment through Wire transfer after receipt of material in good condition.

4) Payment in Foreign Currency for AMC/CMC/Services:

Quarterly/Half yearly after satisfactory completion of work/service duly certified by the H.O.D./P.I. through wire transfer.

E. दंड/ Penalty:

A penalty of 0.5% of the total order value per week shall be levied for the delay subject to a maximum of 5% of the total order value. In case of the non-supply of the item after the acceptance of the work/supply order, EMD will be forfeited by the Institute.

F. अप्रत्याशित घटना/ Force Majeure:

For the purpose of and within the scope of contract by way of indication and not of limitation, the term "Force Mejeure" shall mean acts of nature, strikes, Lockouts, or other industrial disturbances, act of public/enemy, wars, blockades, insurrection, riots, epidemics, landslides, earthquakes, storms lightening, flood, washouts, civil disturbances, explosion and any other similar even not within the control of either party on which by exercise of due care and diligence neither party is able to prevent or overcome.

G. मध्यस्थता/Arbitration:

In the event of any difference or dispute arising out this work, all effort shall be made by both the parties to settle the same amicably. Failing an amicable settlement, the dispute shall be finally settled through arbitration conducted under the Arbitration and Conciliation Act 1996, by arbitrators appointed in accordance with the said Act. The arbitration shall give reasoned and speaking award. All disputes are subject to Dehradun jurisdiction only.

H. बोली अस्वीकृती के लिए मानदंड/ Criteria for bid rejection:

- 1) If not received on or before due date and time.
- 2) If technical bid found without the required undertaking (Annexure-II).

- 3) If it found at a later date that any information given in the bids is incorrect/false then the bid is liable to be disqualified/rejected.
- 4) Canvassing in any form will result in disqualification.
- 5) If "extra, as actual" etc. are mentioned against any of the price components in the bid the bid will be rejected.
- 6) Technical bids should be sealed. More cover, Financial / Commercial bid should also be sealed separately. The sealed envelopes of technical and financial bids should be further put in the main envelop. Main envelop should be sealed. If the financial bid envelop is found opened/unsealed, then bid shall be rejected outrightly.
- 7) If any bidder submitting two/multiple bids for the same of a tender, both/all the bids if that bidder will be outrightly rejected.
- 8) If the firm quotes 'NIL' charges/ consideration, the bid shall be treated as unresponsive and will not be considered.

I. सत्यनिष्ठा संहिता /Code of Integrity

False declarations will be in breach of the Code of Integrity under Rule 175(1)(i)(h) of the General Financial Rules for which a bidder or its successors can be debarred for upto 2 years as per rule 151 (iii) of the General Financial Rules along-with such other actions as may be permissible under law.

Chapter-3: Schedules of requirements

A. टिप्पणी/Note:

- 1) If the Supplier/firm is manufacturer/authorized dealer/sole distributor/of the item, the certificate to this effect should be attached.
- 2) Please note that the firms must submit the compliance Statement in an organized and structured manner in respect of all the specifications as per Annexure-V with the supporting printed catalogue/leaflet of the firm. **Annexure-I, II, III & VI have to be submitted with Technical Bid and Annexure-IV & V with Financial Bid.**
- 3) It may also be noted that the tender bids received after the stipulated date & time or found incomplete and tender bids containing false/incorrect information shall be summarily rejected. The Institute shall not entertain any communication in this regard, whatsoever.
- 4) The firm is required to give its Bank Account details with IFSC for the purpose of making e-payment.
- 5) The Forest Research Institute (FRI) Dehradun reserves the right to reject any quotation wholly or partly without assigning any reason.
- 6) The quantity shown against the item is approximate and may vary as per demand of the Institute at the time placing order.
- 7) The decision of the Institute in all matters relating to eligibility, acceptance, rejection of the bid will be final and binding on the applicants.
- 8) Forest Research Institute (FRI) Dehradun discourages High Sea Sale Purchase. All tenders with High Sea Sale will be rejected.
- 9) The firms should clearly mention (in the financial bid) the Harmonized System of Classification (HS code) defined by the Central Custom and Excise Board (Govt. of India) for the item (s) involving import.
- 10) If the due date happens to be a holiday, then the bid will be accepted till 03.00 PM and opened at 04.00 PM of the next working day followed by the holiday.
- 11) If an agent submits bid on behalf of the Principal/OEM, the same agent can not submit a bid on behalf of another Principal/OEM in the same tender for the same item/product.
- 12) **Indian Agent should submit the certificate from the Foreign Principal that they are the Authorized & Registered Indian Agent.**
- 13) The quantity shown in the tender can be increased or decreased to any extent depending upon the actual requirement. Any corrigendum/addendum/errata in respect of the Tender Notice/ Enquiry letter shall be made available only at our Institute website <http://fri.icfre.gov.in/tender/>. Hence prospective bidders are advised to visit the Institute website regularly.
- 14) Bidders shall be responsible for the correctness of the information provided in the enquiry letter/tender document. If it is found at a later date that any information given in the bid is incorrect/false then the bid is liable to be disqualified/rejected.
- 15) Institute will not be responsible for non-receipt of bid(s) within the stipulated date due to any postal delay or delay for any other reason in transit.
- 16) Before submission of bids, bidders should read the complete enquiry letter/tender document carefully and ensure that the bidders fulfill the eligibility criteria.
- 17) In case the item /product is under rate contract with DGS&D or GeM, then bidder should certify that the rate quoted is not over and above (higher side) the rates under RC with the DGS&D or GeM.
- 18) In case calculation mistakes are found in the financial bid, then calculation shall be made by the Institute on the basis of rate per unit quoted by the bidders will be processed further.
- 19) The Institute reserves the right to rectify any discrepancy of this advertisement, if found later on. In case of any inadvertent mistake in the process which may be detected at any stage even after the issue of purchase order, the Institute reserves the right to modify/withdraw/cancel any communication made to the bidder.

Chapter-4: Specification and allied Technical details

Plz. See Annexure-I, II, III & VI

Chapter-5: Price Schedule (to be utilized by the bidders for quoting their prices)


Plz. See Annexure- V

Chapter 6: Contract Form

Purchase Order is used as Contract Form

Chapter-7: Other Standard Forms, if any to be utilized by the purchaser and bidders

Plz. See Annexure - IV


[PURCHASE OFFICER, FRI]



तकनीकी बोली

(To be given on Company Letter Head)

विस्तृत विवरण अनुपालन रिपोर्ट के साथ प्रस्तुत किया जाना है (अपने आधिकारिक लेटरहेड पर विक्रेता/ फर्म द्वारा प्रस्तुत करने के लिए)
Detailed specification cum compliance report to be submitted with the Technical Bid
(to be submitted by the vender/firm on its official letter head)

(For Detailed Technical Specification Plz refer Annexure-VI)

Sr. No.	Name & Required Specifications of item (s)	Qty.	EMD	Offered specification of item (s) by the firm
1.				
2.				
3.				
4.				
5.				
6.				

	Firm's Profile
1)	Manufacturer/Authorized Agent/Distributor/Dealer/Supplier (Valid certificate must be attached)
2)	E-mail ID & Telephone/Mobile No. Detail of Bank A/C
Note:	(i) "Offered specification" should be filled up properly in the format provided as per Annexure-I
	(ii) Volume & Weight of the item should also be mentioned.
	(iii) The HSN / SAC Code of the item must be mentioned.

Signature:

Name:

Designation:

Seal of the Organisation



उपक्रम/UNDERTAKING
TENDER ACCEPTANCE LETTER
(To be given on Company Letter Head)

Date:

To,

Director
Forest Research Institute
Dehradun-248006

Sub: Acceptance of Terms & Conditions of Tender.

Tender Ref. No X-67/GTI(PP)/SB&RKM/CAMPA/2019-20

Date: 12/02/2020

Agreement/Contract Form

Dear Sir,

- (i) I / We hereby certify that I / we have read the entire terms and conditions of the tender documents from (including all documents like annexure(s), schedule(s), etc.), which form part of the contract agreement and I / we shall abide hereby by the terms / conditions / clauses contained therein.
- (ii) The corrigendum(s) issued from time to time by your department/organization too have also been taken into consideration, while submitting this acceptance letter
- (iii) I / We hereby unconditionally accept the tender conditions of above mentioned tender document(s) / corrigendum(s) in its totality / entirety.
- (iv) **I / We do hereby declare that our Firm has not been blacklisted/ debarred / banned by any Govt. Department/Public sector undertaking.**
- (v) I / We certify that all information furnished by the our Firm is true & correct and in the event that the information is found to be incorrect/untrue or found violated, then your department/ organization shall without giving any notice or reason therefore or summarily reject the bid or terminate the contract, without prejudice to any other rights or remedy including the forfeiture of the full said earnest money deposit absolutely.
- (vi) It is certified that the rate quoted is not over and above (higher side) the rates under RC with the DGS&D or GeM.

Signature:

Name:

Designation:

Seal of the Firm/Agency

CHECK LIST

SI No	(A) Qualifying documents	Yes/No	Page No.
1.	Tender fees for Rs. 590/- (including 18% Service Tax) in shape of crossed D.D in favour of Director, Forest Research Institute, Dehradun or Signed and scanned copy of Exemption certificate like MSME, KVIC, NSIC etc.		
2.	EMD for item No. (1) of Rs. 44,000/- (2) of Rs. 6,000/- item No. (3) of Rs. 24,000/- item No. (4) of Rs. 30,000, item No. (5) of Rs. 10,000/- and item No (6) of Rs.6,000/- in shape of FDR in favour of Director, Forest Research Institute, Dehradun or Signed and scanned copy of Exemption certificate like MSME/ NSIC/KVIC etc.		
3.	Signed and scanned copy of AUTHORIZATION certificate, if applicable.		
4.	Signed and scanned copy of PAN card		
5.	Signed and scanned copy of TENDER ACCEPTANCE LETTER as per tender document Annexure II.		
6.	Signed and Scanned copy of GST certificate, if applicable.		
7.	Signed and Scanned copy of Affidavit of PARTNERSHIP DEED , if applicable.		
8.	Signed and Scanned copy of AUDITED BALANCE SHEET of last two years.		

SI No	(B) Technical Documents	Yes/No	Page No.
1.	Signed and Scanned copy of the self certification only from the local suppliers that the item offered meets the minimum local content (50 % or more Indian manufactured parts are used in the item by cost/value) and shall give the details of the location at which the local value addition is made, wherever applicable.		
2.	Scanned and signed copy of users list and supportive documents from the users for the satisfactory performance and after sales service showing. The firm must enclose the last two supply orders for Reasonability of Rates.		
3.	All specifications should be certified and guaranteed for GENUINENESS & BRAND NEW ITEM and that the model of the equipment is NOT GOING TO BE OBSOLETE in near future (5 years) (Scanned and signed copy of supportive documents from the Principal to be enclosed).		
4.	AVAILABILITY OF SPARES/PARTS for next five years to be ensured (supportive documents from the Principal to be enclosed)		

Note:

1. It is compulsory to fill the above check list with pagination.
2. The document should be submitted/uploaded online in CPP portal in the above order of the check list and place the check list as cover page with pagination of documents.
3. In the absence of any of the qualifying documents, the bid shall be rejected.
4. Tender fee:-The tender form downloaded from above web sites must include **Rs. 590/- (including 18% Service Tax)** by crossed Demand Draft in favour of Director, FRI as tender fee and be **submitted by post/hand to Purchase Officer, Forest Research Institute, Dehradun-248006** on or before the opening date of technical bid, and the scanned copy of the same should be uploaded in the CPP Portal before date-time of submission technical bids.

Signature:

Name:

Designation:

Seal of the Firm/Agency

Mandatory field to fill up by the supplier (enclosed with financial bid only):

Ref. No X-67/GTI(PP)/SB&RKM/CAMPA/2019-20

Date: 12/02/2020

1. Please tick (v) carefully the followings as per requirement:

- A. Customs Duty Exemption Certificate (with following supporting document) []
- (i) GATT Declaration
- (ii) Authority letter for custom Clearance
- (iii) Bank Release Order (BRO)
- (iv)
- (v)
- (vi)
- (vii)
- B. Road Permit / Trade tax Declaration (against Form-32). []
- C. GST Exemption Certificate. []

2. Please fill up the following details:

- A. Complete address of Freight Forwarder, Air lines etc. in whose favor Bank Release Order (BRO) has to be issued:

M/s

- B. Complete address of Custom House Agent (Cha)/ Clearing Agent in whose favor GATT / Authorization for custom clearance has to be issued.

M/s

- C. Complete address of the firm / supplier / CHA to whom original clearance documents will be sent:

M/s

3. Supplier has to submit above details with the financial bid wherever possible and has to send the above detail toand our CHA well ahead in time prior to shipment.

4. Forest Research Institute, Dehradun Custom House Agent (CHA) Address

M/s Cwick (Import Consolidation), Private Limited,
L-II/101, New Mahavir Nagar Extension,
Outer Ring Road, New Delhi-110018,
E-mail: info@cwickimport.com and followup@cwickimport.com
Tel: + 91-09313996082, + 91-11-25996082, + 91-11-25992781

5. Details of Institute:

Import Export Code (IEC code FRI)	0100000011	
AD CODE	02900682870007	Union Bank of India, Main Branch, Rajpur Road, Dehradun-248001

Financial Bid

(To be submitted by the vendor/firm on its official letter head)

Ref. No X-67/GTI(PP)/SB&RKM/CAMPA/2019-20

Dated: 12/02/2020

Rates:						
Sr. No.	Name of item / Specifications/	Qty.	Unit Rate (Rs.)	Amount (Rs.)	GST etc. @.....%	TOTAL (Rs.)
1.	Deep Freezer (-80deg. C)	02				
2.	High throughput horizontal gel electrophoresis systems with power pack	01				
3.	Thermal Cycler (PCR machine)	02				
4.	Real Time PCR or qRT-PCR	01				
5.	Refrigerated Micro centrifuge	01				
6.	Refrigerated Incubator Shaker	01				
Total						
Grand Total Rs.						

- 1) Price of the items should be clearly mentioned if supplied Free of Cost.

Term & Conditions:

Prices (if quoted in Rs.):

FOR FRI, DEHRADUN

1. Prices (if quoted in Foreign Currency) : CIP/CIF New Delhi OR FOB/FCA/Ex-Work

2. Payment:

3. Validity:

4. Delivery period:

5. Other Charges:

6. Warranty:

Onsite 03 Years

7. Rejection:

Equipment if found not as per the required specification would be rejected.

8. Penalty Clause:

As per clause-E of enquiry letter.

In case of the non-supply of the item after the acceptance of the work/supply order, EMD will be forfeited by the Institute.

9. PBG:

5% of the cost price in the form of Bank Guarantee or FDR pledged in favour of "Director, Forest Research Institute, Dehradun" is required at the time of installation of the equipment which should remain valid at least during the warranty period plus sixty days.

It is certified that the rate quoted is not over and above (higher side) the rates under RC with DGS&D or GeM.

Signature:

Name:

Designation:

Seal of the firm/Organisation

Item No. 1 Deep Freezer (-80deg. C)

Deep Freezer (-85° C) : Microprocessor Controlled Touch Screen LCD display Upright Ultra Low Temp. Deep Freezer, 220V, 50Hz.

Specification/Features:**Technical Data:-**

Cabinet Type : Upright

Cooling Type : Direct Cooling, dual compressor

Defrost Mode : Manual

Refrigerant : CFC- Free/ HCFC free

Noise Level : 49dB

Performance:-

Cooling : -85°C/-86° C

Temp. Range : -40°C to -86°C

Controller : Microprocessor, Display : Touch Screen LCD

Electrical Data:-

Power supply : 220-240/50 V/Hz., Power : 1300 W, Electrical Current: 9A. Power supply should be automatically adjusted through built-in stabilizer. The equipment should preferably have energy saving mode to reduces electricity consumption.

Dimensions:-

Capacity : **Between 720-800 Litres**, compact body (all insulations inside)

Functions: The equipment should have indicators for :

High/Low Temperature, Hot Condenser, Power Failure ,
Sensor error, High/Low Battery, High Amb. Temp. And Door Ajar

Accessories:- Castors, Foot , Test Hole : Yes, 2Shelves/Inner Door : 3/4

USB Interface, Remote Alarm, Temp. Recorder, Co2 Back up System , Touch Screen

- LCD display flashing on touch screen Low Noise
 - Cabin design to reduce operation noise
 - Cabinet temperature can reach -86°C with temp uniformity within 4-5°C
 - Inventory Management System (IMS), Layered display for sample storage
 - Cabinet temperature should be displayed on the front panel
 - Automatic alarm should triggered when temperature exceeds allowable limits
 - Modular Design for Refrigeration system and Electrical Control System, Strong Cooling
- Satisfactory performance certificate for the quoted model from at least five users where the same model has been installed should be submitted

Credibility: The vendor should demonstrate a record of successful installations in similar institutions/research centers across India. A comprehensive list of prior installations should be provided. Two years comprehensive onsite warranty and two years additional Annual Maintenance Contract. Complete with all accessories and original catalogues

Item No. 2 High throughput horizontal gel electrophoresis systems with power pack

Specification


a. High Throughput vertical gel assembly with power pack

- High-Throughput with the plate size of 50cm (w) x 22cm (l) capable of running 100 PCR samples each side (200 total). Dual sided, high-throughput, adjustable height electrophoresis unit with guide rods and rotating/leveling base
- Platinum Electrodes
- Sample wells are multi-channel pipettor compatible
- The gel can be re-used up to six times, significantly reducing cost per sample
- Low fluorescence glass to allows direct UV scanning of ETBR stained gels without removing the gel from the plate sandwich
- The system should have quick and easy gel casting mechanism
- A divided upper reservoir to run identical or independent buffer conditions on either side as well as allows to run just one side of the unit if needed.
- The system should be compatible for Simple Sequence Repeats (SSR) Genotyping, for linkage map construction (micro-satellite markers), Mutation Detection DNA Sequencing, Mini-satellites, Differential Display, DNA Fingerprinting/Foot-printing, R Nase Protection, SDS-PAGE
- Accessory included with the unit: 2 Combs (1.5mm x 100 wells), 2 Sets of Spacer with Gel Wrap (1.5mm x 22 cm), 2 Bar clamps, 2 Sets glass plates (back plate borosilicate) 50 cm (w) x 22 cm (h); Gel sealing tape, Temperature strips, Plate separators, White spring clamps, Guide rods and Rotating base, Safety covers with power leads. Conversion plate for running one gel at a time
- Compatible powerpack - High Voltage Programmable Power Supply, 0-3000V, 50/60Hz, 230V

b. Horizontal electrophoresis system

- High-Throughput Electrophoresis system with flexible gel sizes, along with complete accessories such as gel tank, gel trays, gel caster, gel casting gates, combs, leveling bubble etc. including power packs along with safety cable.
- Required units with following gel tray sizes (approx.): 15cm (w) x 10cm (l), 26cm (w) x 14cm (l), 26cm (w) x 26cm (l) and 26cm (w) x 40cm (l).
- Units with different sample throughput varying from < 20, < 60, < 96, < 120 or more.
- Suitable Gel tanks to accommodate gel of above mentioned sizes and with the capacity of base buffer volume ~ 500ml to 3000ml.
- Combs of varying No. of wells corresponding to above gel trays should be quoted.
- Universal power packs with programmable output range compatible with required electrophoresis units.

The complete system should also have following additional features:

- Provision of buffer recirculation ports to allow maintenance of uniform pH and ionic strength and reduce frequency of buffer changes necessary during long runs.
 - Gel Tray should be made up of UV transparent acrylic and it should have configuration allowing single run or multiple simultaneous short runs.
 - Combs with adjustable height and comb holders for keeping the combs perfectly straight during gel casting with hot agarose. Also the combs should allow multi-channel pipetting for accuracy and speed.
 - Adequate safety measures to prevent electric shocks, leaking etc. such as as per the standards.
 - Provision of keeping constant temperature during runs.
 - Color-coded, labeled electrodes and labeled base for correct positioning of the lid on the base
- 

Item No. 3 Thermal Cycler (PCR machine)

Specification

1. Gradient PCR with silver block, with sample capacity of 96 x 0.2 ml PCR tubes, or 1 micro test plate 8 x 12 (with SBS standard), or 96x0.2ml PCR plate skirted, semi skirted, or non-skirted.
2. Instrument should have a temperature range of 4-99° C with temperature accuracy of + 0.2°C to + 0.3°C; block homogeneity of + 0.3°C to + 0.4°C.
3. Instrument should have a ramp rate of at least 6 ° C/s and 4.5° C/s or more for cooling-heating respectively, and adjustable ramp rate to meet critical amplification conditions.
4. The ramp rates should be identical in both gradient and normal operation.
5. The heated lid should have a temperature range of 37-110°C, must be able to accommodate both flat and dome capped tubes, should allow the user to use a reaction volume as low as 5ul, ensuring minimal sample evaporation throughout the block..
6. The instrument should have a gradient range of 30-99°C and the range of gradient span should be 1°C –24°C.
7. The instrument should have a high resolution, large display for programs.
8. Instrument should be equipped with intuitive graphical interface for rapid input of protocols, and easy file management.
9. Instrument should offer the capacity to store minimum of 700 or more programs
10. The instrument should have the provision for connecting 4-5 more blocks (96 or 384 well), if required at later stages.
11. The instrument should have two or more USB ports to attach mouse and/or memory stick.
12. The provision for networking 20 or more units through system software, with one computer, should be provided.
13. Instrument should have universal power failure protection, with automatic restart after power failure up to 90 min or more.
14. Instrument should have temperature control modes: Fast, Standard and Safe
15. Satisfactory performance certificate for the quoted model from at least five users where the same model has been installed should be submitted
16. Credibility: The vendor should demonstrate a record of successful installations in similar institutions/research centers across India. A comprehensive list of prior installations should be provided.

Item No. 4 Real Time PCR or qRT-PCR

Specification


Specifications for Real time PCR or qRT-PCR		Score (Total)
Reaction volume	96-well, 0.2 mL block: 10–100 µL upgradable upto 384-well or more	4
Excitation source	Bright-white LED/ different LEDs	3
Optical detection	6 filters , CMOS camera	6
Excitation/detection range	450–680 nm/500–730 nm	5
Temperature range	4–100°C	5
Maximum block /sample ramp rate	5 -7°C/sec	8
Average sample ramp rate	3 to 3.6°C/sec	4
Temperature uniformity	0.5°C	5
Temperature accuracy	0.25°C	5
Heating and cooling method	Peltier	6
Independent temperature zones	Temperature cycling for gradient control (linear gradient / 6 zones, 5°C zone to zone)	7
Chemistries	Both fast and standard	5
Run times	At least 30 min (fast mode)	4
Validated reaction volumes	10–100 µL	3
Factory-calibrated dyes	FAM™, SYBR™, VIC™, ABY™, NED™, TAMRA™, JUN™, ROX™, Mustang Purple™, and Cy®5 dyes HEX, Texas Red	6
Multiplex	5-plex with 1 passive reference; 6-plex with no passive reference ; or more than 5-plex	5
Sensitivity	1 copy; detect differences as small as 1.5-fold in target quantities in singleplex	5
Dynamic range	9 logarithmic units of linear dynamic range	5
Security and auditing features	21 CFR Part 11 / MIQE RDML compliance	7
Suitable PC with UPS along with Consumable	<ul style="list-style-type: none"> • PC with minimum i5 processor or equivalent with original window and software for running Real time PCR application, 2 KV UPS, • Consumables to be included for immediate use for 200 reactions 	2

Item No. 5 Refrigerated Micro centrifuge

Specification

- Minimum capacity up to $48 \times 1.5/2.0$ mL.
- Minimum Speed up to $25,000 \times g$ (16,220 rpm) with accuracy of 10 rpm
- Temperature range from -10°C to $+40^\circ\text{C}$
- Compact in size and able to accommodate a fixed-angle rotor for 1.5/2.0 ml micro tubes
- Microprocessor controlled Touch Screen or alphanumeric LCD Screen for display of set & run parameters i.e. Speed, RCF, Temperature, Run Time, etc.
- Maintenance-free, noiseless and brushless drive motor
- Automatic rotor identification system
- Quick Lock-system for quick rotor lid opening and closing
- Fast Temp program for fast pre-cooling and maximum temperature accuracy inside the rotor
- Dimensions (W \times D \times H) should be around $319 \times 540 \times 254$ mm.
- Power supply: 220-240V AC, 50/60 Hz
- System should have International certification like IVD conformity.
- System should be quoted with suitable voltage stabilizer.
- One year warranty with spares
- Company should have good No. of installation and user list also attached with technical bid.

Required Accessories

- Suitable autoclavable 48-place fixed angle rotor for 1.5/2.0 ml micro tubes.
 - Rotor accessories and suitable adaptors
 - Suitable voltage stabilizer
- 

Item No. 6 Refrigerated Incubator Shaker

Specification

1. Compact tabletop refrigerated shaker with minimal footprint and speed ranged from 30 to 350 RPM with accuracy of maximum ± 2 .
 2. To ensure non-deviated amplitude of minimum 25 mm for uniform growth of culture, machine should have triple eccentric counter balance mechanism and mounted on permanently lubricated bearings.
 3. Machine should have programmable Digital LED display and microprocessor based PID controller for temperature.
 4. Machine should have temperature range from 7 to 70°C with accuracy of $\pm 0.1^\circ\text{C}$.
 5. Should use finned tubular air heater for good heat transfer.
 6. Temperature control through soft keypad with visual and / or audio alarm for temperature over / under shoot. Thermostat cut-off for temperature over shoot.
 7. Machine should have universal platform of size 420x420 mm to accommodate a minimum of 36 Nos. of 100 ml (or) 25 Nos. of 250 ml. (or) 16 Nos. of 500 ml. (or) 9 Nos. of 1000 ml. (or) 4 Nos. of 2000 ml flasks.
 8. The platform should be provided with universal tray with suitable clamps to accommodate different types or sizes of labwares (flasks or culture tubes). Clamp must be made of single piece stainless steel material and leaves tied with three springs.
 9. Option for adjustable stationary tray for incubation should be available
 10. High-grade insulation on all sides for uniform temperature maintenance and material of construction should be MS powder coated steel.
 11. Equipment should be future stackable using a stack rack.
 12. Should have options for future up-gradation of illumination, UV light for decontamination.
 13. Only standard models should be quoted and custom-made products will not be preferred.
 14. Machine should be quoted with minimum two year warranty
 15. Company should have minimum 20 installations of the quoted model across India in other government institutions or universities like IIT's, CSIR, ICAR, ICMR labs and installation certificates must be attached along with quotation.
- 